
Midmark M11 - v1.0.0
Total Cycles: 1488

Name: RITTER 1
Sterilizer ID: V2178198
12/29/2021 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.6 F 0.0
2:00 203.9 F 3.5
4:00 202.0 F 8.6
6:00 224.2 F 9.4
8:00 231.5 F 9.7
10:00 235.4 F 8.5
12:00 243.9 F 13.2
14:00 243.2 F 11.9
16:00 255.9 F 18.6
18:00 267.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.4 F 28.9
2:00 271.5 F 28.9
3:00 272.2 F 28.8
4:00 271.4 F 28.8

Min 271.2 F 28.4
Max 272.5 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.4 F 0.0
5:00 225.0 F 0.0
10:00 232.6 F 0.0
15:00 228.6 F 0.0
20:00 229.3 F 0.0
25:00 229.5 F 0.0
30:00 227.5 F 0.0

Drying Complete

FILLING: 2:06
HEATING: 19:25
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

12/29/2021 02:16 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____