
Midmark M11 - v1.0.0
Total Cycles: 3665

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2023 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.7 F 0.0
2:00 168.3 F 3.1
4:00 176.8 F 6.8
6:00 198.4 F 11.0
8:00 218.4 F 4.6
10:00 231.8 F 9.2
12:00 239.5 F 11.0
14:00 239.3 F 10.8
16:00 235.7 F 8.5
18:00 249.8 F 15.1
20:00 261.1 F 21.5
22:00 271.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.9
2:00 271.6 F 28.6
3:00 272.2 F 28.8
4:00 271.4 F 28.5

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.7 F 0.0
5:00 236.0 F 0.0
10:00 226.5 F 0.0
15:00 220.8 F 0.0
20:00 211.9 F 0.0
25:00 219.6 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:01:18
CYCLE COMPLETE

08/16/2023 04:25 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____