
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006151
12:50 17/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:53
STERILIZING 3:54
136.4 C 221kPa 3:54
136.2 C 224kPa 4:24
136.4 C 223kPa 4:54
136.4 C 225kPa 5:24
136.1 C 220kPa 5:54
136.1 C 219kPa 6:24
136.4 C 220kPa 6:54
136.4 C 223kPa 7:24
136.4 C 223kPa 7:24
VENTING 7:24
AIR DRYING 8:15
CYCLE COMPLETE 16:22

Digital Signature #
E66201693EC2CC26
