
Midmark M11 - v1.0.0
Total Cycles: 1760

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2022 10:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.0 F 0.0
2:00 152.4 F 1.7
4:00 172.4 F 4.6
6:00 186.8 F 8.7
8:00 208.8 F 11.2
10:00 221.1 F 7.2
12:00 233.6 F 12.1
14:00 234.6 F 8.8
16:00 240.5 F 10.2
18:00 246.3 F 13.3
20:00 258.3 F 19.9
22:00 269.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.6 F 29.1
2:00 272.1 F 28.4
3:00 271.5 F 28.5
4:00 272.6 F 29.0

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 223.4 F 0.0
10:00 230.2 F 0.0
15:00 231.2 F 0.0
20:00 228.3 F 0.0
25:00 228.2 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:04
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

03/29/2022 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____