
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000456
8:14 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 4:24
136.0 C 220kPa 4:24
136.2 C 218kPa 4:54
136.4 C 224kPa 5:24
136.2 C 223kPa 5:54
136.4 C 221kPa 6:24
135.9 C 221kPa 6:54
136.2 C 219kPa 7:24
136.4 C 225kPa 7:54
136.4 C 225kPa 7:54
VENTING 7:54
AIR DRYING 8:45
CYCLE COMPLETE 9:01

Digital Signature #
00C85BC267F09C7E
