
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001026
16:26 06/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 2:00
STERILIZING 3:01
136.4 C 225kPa 3:01
136.4 C 224kPa 3:31
136.2 C 223kPa 4:01
136.4 C 226kPa 4:31
136.2 C 224kPa 5:01
136.4 C 226kPa 5:31
136.4 C 226kPa 6:01
136.4 C 226kPa 6:31
136.4 C 226kPa 6:31
VENTING 6:31
AIR DRYING 7:22
CYCLE COMPLETE 22:22

Digital Signature #
93CBA4CC253A227A
