
Midmark M11 - v1.0.0
Total Cycles: 2304

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2022 11:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.0 F 0.0
2:00 153.0 F 2.5
4:00 172.5 F 7.1
6:00 194.8 F 12.5
8:00 215.1 F 6.5
10:00 230.1 F 11.7
12:00 231.7 F 8.1
14:00 242.1 F 11.6
16:00 243.7 F 12.6
18:00 256.4 F 19.2
20:00 267.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.7 F 29.5
2:00 272.1 F 28.9
3:00 271.9 F 29.0
4:00 272.7 F 29.5

Min 271.6 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 223.1 F 0.0
10:00 225.0 F 0.0
15:00 225.3 F 0.0
20:00 227.7 F 0.0
25:00 229.2 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:20
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:04
CYCLE COMPLETE

09/08/2022 12:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____