
Midmark M11 - v1.0.5
Total Cycles: 3085

Name: RITTER2
Sterilizer ID: V2313228
11/06/2023 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.5 F 0.0
2:00 85.4 F 0.0
4:00 93.8 F 0.2
6:00 104.1 F 1.5
8:00 135.7 F 4.3
10:00 166.6 F 7.2
12:00 202.1 F 10.5
14:00 222.5 F 5.1
16:00 223.7 F 6.9
18:00 232.4 F 10.7
20:00 232.8 F 7.9
22:00 239.4 F 11.4
24:00 236.5 F 9.0
26:00 249.1 F 15.1
28:00 260.0 F 21.2
30:00 270.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.3 F 29.7
2:00 271.5 F 29.1
3:00 272.3 F 29.4
4:00 272.4 F 29.8

Min 270.9 F 28.6
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.2 F 0.0
5:00 216.7 F 0.0
10:00 230.5 F 0.0
15:00 230.1 F 0.0
20:00 225.0 F 0.0
25:00 227.1 F 0.0
30:00 231.3 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:47
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:08:52
CYCLE COMPLETE

11/06/2023 10:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____