
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000249
15:11 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:54
STERILIZING 5:16
136.2 C 230kPa 5:16
136.4 C 233kPa 5:46
136.4 C 231kPa 6:16
136.2 C 233kPa 6:46
136.4 C 233kPa 7:16
136.4 C 233kPa 7:46
136.3 C 233kPa 8:16
136.3 C 233kPa 8:46
136.3 C 233kPa 8:46
VENTING 8:46
AIR DRYING 9:50
CYCLE COMPLETE 10:31

Digital Signature #
A9FFAD4717C20D88
