
Midmark M11 - v1.0.5
Total Cycles: 1041

Name: RITTER2
Sterilizer ID: V2313228
04/27/2022 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.9 F 0.0
2:00 197.4 F 3.1
4:00 207.3 F 8.7
6:00 229.8 F 9.1
8:00 235.0 F 11.1
10:00 235.5 F 9.0
12:00 239.1 F 9.7
14:00 248.7 F 15.1
16:00 261.0 F 22.2
18:00 272.5 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.4 F 29.7
2:00 271.1 F 29.0
3:00 272.2 F 29.9
4:00 272.0 F 29.3

Min 271.0 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 213.2 F 0.0
10:00 221.6 F 0.0
15:00 224.2 F 0.0
20:00 228.2 F 0.0
25:00 228.7 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 18:27
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

04/27/2022 10:35 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____