
Midmark M11 - v1.0.5
Total Cycles: 1268

Name: RITTER2
Sterilizer ID: V2313228
07/07/2022 12:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.1 F 0.0
2:00 156.3 F 2.6
4:00 168.7 F 6.9
6:00 194.1 F 12.0
8:00 214.9 F 6.0
10:00 228.8 F 10.9
12:00 229.3 F 7.4
14:00 243.0 F 13.5
16:00 242.4 F 12.3
18:00 254.7 F 18.8
20:00 265.9 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.7 F 30.2
2:00 272.6 F 30.0
3:00 272.1 F 29.5
4:00 271.4 F 29.2

Min 271.4 F 29.2
Max 273.0 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.7 F 0.0
5:00 218.3 F 0.0
10:00 219.6 F 0.0
15:00 217.9 F 0.0
20:00 214.8 F 0.0
25:00 211.5 F 0.0
30:00 202.5 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

07/07/2022 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____