
Midmark M11 - v1.0.0
Total Cycles: 2327

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2022 02:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.3 F 0.0
2:00 196.6 F 4.4
4:00 212.0 F 9.3
6:00 223.3 F 6.3
8:00 233.8 F 10.5
10:00 234.0 F 7.3
12:00 242.5 F 11.4
14:00 246.2 F 13.4
16:00 259.2 F 20.4
18:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.3 F 29.1
2:00 271.9 F 28.6
3:00 271.6 F 28.7
4:00 272.4 F 29.0

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 216.6 F 0.0
5:00 237.1 F 0.0
10:00 229.1 F 0.0
15:00 228.1 F 0.0
20:00 226.7 F 0.0
25:00 231.2 F 0.0
30:00 227.3 F 0.0

Drying Complete

FILLING: 2:09
HEATING: 18:38
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

09/14/2022 03:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____