

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000970  
8:28 14/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:55  
PRESSURIZING 3:58  
STERILIZING 5:04  
136.0 C 220kPa 5:04  
136.4 C 224kPa 5:34  
136.1 C 221kPa 6:04  
136.3 C 222kPa 6:34  
136.3 C 222kPa 7:04  
136.3 C 224kPa 7:34  
136.4 C 222kPa 8:04  
136.4 C 224kPa 8:34  
136.4 C 224kPa 8:34  
VENTING 8:34  
AIR DRYING 9:27  
CYCLE COMPLETE 9:46

-----  
Digital Signature #  
66ADC43DA8A1CFB5  
-----