
Midmark M11 - v1.0.0
Total Cycles: 1930

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2022 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.4 F 0.0
2:00 184.6 F 4.2
4:00 190.8 F 8.8
6:00 207.5 F 12.2
8:00 219.0 F 7.9
10:00 232.9 F 13.3
12:00 234.7 F 10.3
14:00 237.5 F 7.9
16:00 248.0 F 14.9
18:00 260.3 F 21.6
20:00 271.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.5
2:00 272.0 F 28.9
3:00 272.1 F 29.3
4:00 272.8 F 29.5

Min 271.5 F 28.7
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 221.4 F 0.0
10:00 230.3 F 0.0
15:00 228.6 F 0.0
20:00 227.5 F 0.0
25:00 228.0 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 20:43
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

05/11/2022 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____