
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000656
15:06 19/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:19
136.0 C 220kPa 4:19
136.2 C 223kPa 4:49
136.4 C 225kPa 5:19
136.4 C 225kPa 5:49
136.5 C 225kPa 6:19
136.5 C 222kPa 6:49
136.4 C 225kPa 7:19
136.4 C 225kPa 7:49
136.4 C 225kPa 7:49
VENTING 7:49
AIR DRYING 8:46
CYCLE COMPLETE 8:59

Digital Signature #
DC07E5DCDAF61F12
