
Midmark M11 - v1.0.0
Total Cycles: 3713

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2023 03:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.4 F 0.0
2:00 161.2 F 3.4
4:00 182.1 F 7.4
6:00 203.5 F 12.2
8:00 220.6 F 5.9
10:00 232.1 F 10.6
12:00 230.7 F 7.0
14:00 243.5 F 12.6
16:00 240.9 F 10.7
18:00 253.8 F 17.1
20:00 264.8 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.9 F 29.0
2:00 272.4 F 29.2
3:00 271.6 F 28.5
4:00 272.2 F 28.9

Min 270.9 F 28.1
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 219.1 F 0.0
10:00 217.5 F 0.0
15:00 209.7 F 0.0
20:00 205.8 F 0.0
25:00 209.3 F 0.0
30:00 205.1 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:54
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

08/31/2023 04:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____