

-----  
Midmark M11 - v1.0.5  
Total Cycles: 608

Name: RITTER2  
Sterilizer ID: V2313228  
12/07/2021 11:55 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.0 F 0.0  
2:00 145.6 F 0.8  
4:00 166.1 F 3.9  
6:00 190.4 F 8.9  
8:00 213.8 F 6.3  
10:00 229.4 F 10.1  
12:00 232.9 F 7.9  
14:00 241.2 F 10.9  
16:00 247.4 F 14.1  
18:00 260.2 F 21.3  
20:00 271.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 272.3 F 28.9  
2:00 272.9 F 29.3  
3:00 272.7 F 29.6  
  
Min 271.8 F 28.6  
Max 273.0 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 169.3 F 0.0  
10:00 164.3 F 0.0  
15:00 159.9 F 0.0  
20:00 154.5 F 0.0  
Dry Cycle Aborted

FILLING: 2:25  
HEATING: 20:32  
STERILIZING: 3:00  
VENTING: 2:45  
DRYING: 24:01  
TOTAL CYCLE: 00:52:43

12/08/2021 07:26 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02