
Midmark M11 - v1.0.0
Total Cycles: 4502

Name: RITTER 1
Sterilizer ID: V2178198
04/29/2024 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.2 F 0.0
2:00 85.7 F 0.1
4:00 97.2 F 1.6
6:00 126.0 F 3.8
8:00 153.3 F 6.3
10:00 171.7 F 9.4
12:00 199.3 F 12.6
14:00 207.8 F 4.8
16:00 219.3 F 8.1
18:00 229.2 F 11.8
20:00 228.4 F 7.3
22:00 240.2 F 12.2
24:00 238.0 F 9.8
26:00 250.1 F 15.4
28:00 260.3 F 21.1
30:00 270.1 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 272.5 F 29.1
2:00 272.2 F 29.4
3:00 272.0 F 28.9
4:00 272.6 F 29.5

Min 271.6 F 28.7
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 196.2 F 0.0
10:00 202.8 F 0.0
15:00 203.8 F 0.0
20:00 220.8 F 0.0
25:00 227.8 F 0.0
30:00 234.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:41
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:10:23
CYCLE COMPLETE

04/29/2024 02:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____