
Midmark M11 - v1.0.0
Total Cycles: 1782

Name: RITTER 1
Sterilizer ID: V2178198
04/01/2022 11:17 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	143.5 F	0.0
2:00	140.3 F	1.8
4:00	154.2 F	4.6
6:00	182.3 F	8.6
8:00	203.7 F	7.3
10:00	224.3 F	8.1
12:00	239.1 F	13.5
14:00	238.5 F	10.5
16:00	234.9 F	8.3
18:00	250.1 F	15.3
20:00	257.2 F	19.2

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 258F
CHAMBER PRESS 19.6PSI
** ITEMS NOT STERILE **

04/01/2022 11:39 AM