
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001161
13:26 10/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:22
STERILIZING 3:25
136.5 C 220kPa 3:25
136.4 C 225kPa 3:55
136.3 C 224kPa 4:25
136.4 C 223kPa 4:55
136.4 C 220kPa 5:25
136.1 C 223kPa 5:55
136.3 C 223kPa 6:25
136.2 C 223kPa 6:55
136.2 C 223kPa 6:55
VENTING 6:55
AIR DRYING 7:44
CYCLE COMPLETE 7:45

Digital Signature #
200600390825390A
