

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4714

Name: RITTER 1  
Sterilizer ID: V2178198  
07/11/2024 08:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.0 F 0.0  
2:00 71.4 F 0.2  
4:00 91.0 F 1.2  
6:00 124.5 F 3.5  
8:00 157.4 F 6.5  
10:00 180.8 F 9.9  
12:00 199.9 F 13.4  
14:00 212.9 F 4.1  
16:00 221.7 F 7.3  
18:00 230.1 F 10.8  
20:00 230.1 F 7.3  
22:00 239.4 F 10.6  
24:00 234.3 F 7.7  
26:00 247.4 F 13.8  
28:00 258.4 F 19.8  
30:00 268.4 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.3 F 28.5  
2:00 272.0 F 28.9  
3:00 271.5 F 28.1  
4:00 272.2 F 28.9

Min 271.1 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 234.9 F 0.0  
10:00 220.1 F 0.0  
15:00 225.9 F 0.0  
20:00 194.3 F 0.0  
25:00 195.0 F 0.0  
30:00 170.8 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:14  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:09:28  
CYCLE COMPLETE

07/11/2024 09:24 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_