
Midmark M11 - v1.0.0
Total Cycles: 2157

Name: RITTER 1
Sterilizer ID: V2178198
07/20/2022 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.3 F 0.0
2:00 164.0 F 3.3
4:00 179.1 F 7.2
6:00 209.0 F 11.7
8:00 217.0 F 5.9
10:00 231.7 F 10.9
12:00 230.4 F 6.7
14:00 242.4 F 13.0
16:00 241.0 F 11.1
18:00 252.9 F 17.5
20:00 264.6 F 24.1
22:00 271.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.3 F 29.6
2:00 272.1 F 29.2
3:00 271.2 F 28.6
4:00 271.7 F 29.3

Min 271.1 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 226.3 F 0.0
10:00 235.2 F 0.0
15:00 230.0 F 0.0
20:00 231.1 F 0.0
25:00 228.6 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

07/20/2022 02:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____