

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2607

Name: RITTER 1  
Sterilizer ID: V2178198  
11/23/2022 04:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
2:00 197.1 F 5.1  
4:00 207.4 F 10.8  
6:00 223.5 F 6.5  
8:00 235.9 F 12.8  
10:00 239.1 F 10.8  
12:00 238.1 F 9.6  
14:00 253.0 F 17.1  
16:00 265.5 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 272.7 F 29.4  
2:00 272.2 F 28.8  
3:00 271.7 F 28.7  
4:00 272.8 F 29.5  
  
Min 271.6 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 233.3 F 0.0  
10:00 230.8 F 0.0  
15:00 227.1 F 0.0  
20:00 228.1 F 0.0  
25:00 229.9 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 2:23  
HEATING: 17:42  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:18  
CYCLE COMPLETE

11/23/2022 05:10 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_