

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001326  
11:28 28/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:26  
PRESSURIZING 2:31  
STERILIZING 3:39  
136.0 C 220kPa 3:39  
136.3 C 223kPa 4:09  
136.4 C 225kPa 4:39  
136.3 C 221kPa 5:09  
136.2 C 224kPa 5:39  
136.4 C 225kPa 6:09  
136.2 C 222kPa 6:39  
136.2 C 222kPa 7:09  
136.2 C 222kPa 7:09  
VENTING 7:09  
AIR DRYING 8:05  
CYCLE COMPLETE 23:05

-----  
Digital Signature #  
D0F2233073C2371E  
-----