
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000608
9:27 27/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:35
STERILIZING 4:43
136.0 C 228kPa 4:43
136.4 C 231kPa 5:13
136.2 C 230kPa 5:43
136.4 C 230kPa 6:13
136.3 C 230kPa 6:43
136.4 C 232kPa 7:13
136.1 C 231kPa 7:43
136.4 C 232kPa 8:13
136.4 C 232kPa 8:13
VENTING 8:13
AIR DRYING 9:08
CYCLE COMPLETE 24:08

Digital Signature #
CDAF6D8F1F614A83
