
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007144
8:23 04/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:45
STERILIZING 3:46
136.0 C 220kPa 3:46
136.2 C 223kPa 4:17
136.3 C 224kPa 4:47
136.4 C 223kPa 5:17
136.3 C 221kPa 5:47
136.4 C 221kPa 6:17
135.8 C 218kPa 6:47
136.4 C 223kPa 7:17
136.4 C 223kPa 7:17
VENTING 7:17
AIR DRYING 8:06
CYCLE COMPLETE 8:27

Digital Signature #
5334BFAD4DDAD3A4
