
Midmark M11 - v1.0.0
Total Cycles: 1028

Name: RITTER 1
Sterilizer ID: V2178198
07/19/2021 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.9 F 0.0
2:00 157.5 F 1.5
4:00 165.2 F 4.3
6:00 189.6 F 8.5
8:00 206.9 F 13.3
10:00 216.8 F 8.7
12:00 229.6 F 11.1
14:00 238.6 F 11.0
16:00 236.3 F 8.7
18:00 249.8 F 15.0
20:00 260.6 F 21.4
22:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 271.7 F 29.2
2:00 271.7 F 28.9
3:00 271.2 F 29.0
4:00 271.7 F 29.3

Min 271.0 F 28.8
Max 272.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 230.9 F 0.0
10:00 227.5 F 0.0
15:00 227.7 F 0.0
20:00 211.3 F 0.0
25:00 215.9 F 0.0
30:00 205.5 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 23:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:31
CYCLE COMPLETE

07/19/2021 12:23 PM
ACCEPTED: X
REJECTED:
Operator: U01