
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006916
8:18 08/02/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:52
STERILIZING 3:55
136.4 C 222kPa 3:55
136.4 C 221kPa 4:25
136.2 C 223kPa 4:55
136.2 C 220kPa 5:25
136.4 C 222kPa 5:55
136.0 C 220kPa 6:25
136.4 C 220kPa 6:55
136.3 C 223kPa 7:25
136.3 C 223kPa 7:25
VENTING 7:25
AIR DRYING 8:16
CYCLE COMPLETE 8:43

Digital Signature #
581AB1F282D0D30E
