

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4363

Name: RITTER2  
Sterilizer ID: V2313228  
02/20/2025 12:18 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.9 F 0.0  
2:00 130.1 F 1.2  
4:00 131.8 F 3.5  
6:00 148.5 F 6.8  
8:00 179.1 F 10.4  
10:00 200.9 F 5.5  
12:00 213.0 F 6.9  
14:00 223.7 F 11.0  
16:00 229.6 F 7.4  
18:00 239.5 F 12.3  
20:00 238.2 F 10.5  
22:00 249.4 F 16.2  
24:00 259.8 F 22.2  
26:00 269.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 30.1  
1:00 272.6 F 30.7  
2:00 272.1 F 30.0  
3:00 272.7 F 30.7  
4:00 272.2 F 30.0

Min 271.9 F 29.9  
Max 273.1 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.7 F 0.0  
5:00 203.2 F 0.0  
10:00 208.1 F 0.0  
15:00 200.2 F 0.0  
20:00 194.6 F 0.0  
25:00 186.5 F 0.0  
30:00 180.2 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 26:54  
STERILIZING: 4:00  
VENTING: 3:56  
DRYING: 30:00  
TOTAL CYCLE: 01:05:45  
CYCLE COMPLETE

02/20/2025 01:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_