
Midmark M11 - v1.0.0
Total Cycles: 2783

Name: RITTER 1
Sterilizer ID: V2178198
01/17/2023 10:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 195.0 F 2.3
4:00 199.3 F 6.2
6:00 209.2 F 11.5
8:00 220.3 F 5.8
10:00 234.2 F 11.1
12:00 233.7 F 8.1
14:00 242.8 F 11.5
16:00 245.1 F 12.8
18:00 257.4 F 19.3
20:00 268.4 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 29.0
2:00 271.5 F 28.2
3:00 272.2 F 29.1
4:00 272.3 F 28.8

Min 271.2 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 232.0 F 0.0
10:00 227.9 F 0.0
15:00 230.4 F 0.0
20:00 228.5 F 0.0
25:00 229.8 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:15
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:40
CYCLE COMPLETE

01/17/2023 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____