
Midmark M11 - v1.0.5
Total Cycles: 1730

Name: RITTER2
Sterilizer ID: V2313228
11/16/2022 08:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.2 F 0.0
2:00 79.3 F 0.0
4:00 91.6 F 0.0
6:00 111.6 F 0.0
8:00 154.2 F 0.4
10:00 185.9 F 0.8
12:00 203.3 F 3.6
14:00 218.6 F 7.0
16:00 228.9 F 10.7
18:00 229.2 F 7.7
20:00 238.0 F 11.3
22:00 235.5 F 9.0
24:00 240.0 F 10.2
26:00 247.7 F 14.3
28:00 259.6 F 20.9
30:00 270.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.5 F 29.2
2:00 272.0 F 29.2
3:00 272.2 F 29.0
4:00 272.2 F 29.5

Min 271.6 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 202.7 F 0.0
10:00 214.2 F 0.0
15:00 210.4 F 0.0
20:00 207.2 F 0.0
25:00 206.2 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 30:50
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:50
CYCLE COMPLETE

11/16/2022 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____