
Midmark M11 - v1.0.0
Total Cycles: 5101

Name: RITTER 1
Sterilizer ID: V2178198
11/18/2024 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 68.6 F 0.0
4:00 85.6 F 0.0
6:00 120.5 F 0.0
8:00 151.7 F 0.2
10:00 176.8 F 0.7
12:00 193.4 F 1.2
14:00 211.4 F 4.5
16:00 221.7 F 7.4
18:00 231.1 F 10.7
20:00 230.7 F 7.7
22:00 238.4 F 10.5
24:00 235.5 F 8.2
26:00 246.4 F 13.5
28:00 243.8 F 12.0
30:00 256.1 F 18.4
32:00 267.2 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.1
1:00 272.0 F 28.6
2:00 272.3 F 28.4
3:00 271.5 F 28.2
4:00 272.7 F 28.8

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 208.8 F 0.0
10:00 187.8 F 0.0
15:00 184.9 F 0.0
20:00 186.7 F 0.0
25:00 200.2 F 0.0
30:00 208.0 F 0.0

Drying Complete

FILLING: 0:43
HEATING: 33:21
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:11:33
CYCLE COMPLETE

11/18/2024 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____