
Midmark M11 - v1.0.0
Total Cycles: 1071

Name: RITTER 1
Sterilizer ID: V2178198
08/04/2021 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.3 F 0.0
2:00 110.7 F 0.0
4:00 114.9 F 0.1
6:00 149.1 F 0.4
8:00 175.5 F 2.9
10:00 193.6 F 5.5
12:00 207.0 F 8.6
14:00 218.8 F 12.1
16:00 222.9 F 5.6
18:00 235.3 F 10.2
20:00 236.7 F 8.9
22:00 245.0 F 13.0
24:00 243.3 F 11.6
26:00 255.2 F 17.9
28:00 265.9 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.7
2:00 271.6 F 28.0
3:00 272.7 F 28.9
4:00 272.2 F 28.3

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 211.9 F 0.0
10:00 219.3 F 0.0
15:00 216.3 F 0.0
20:00 199.2 F 0.0
25:00 222.8 F 0.0
30:00 216.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 29:46
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:08:06
CYCLE COMPLETE

08/04/2021 02:36 PM
ACCEPTED: X
REJECTED:
Operator: U01