
Midmark M11 - v1.0.0
Total Cycles: 5290

Name: RITTER 1
Sterilizer ID: V2178198
01/29/2025 10:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 71.5 F 0.0
4:00 86.6 F 0.1
6:00 125.5 F 1.6
8:00 155.2 F 4.1
10:00 178.2 F 6.9
12:00 196.9 F 10.0
14:00 205.2 F 10.7
16:00 217.7 F 5.0
18:00 228.7 F 8.4
20:00 237.6 F 12.3
22:00 231.7 F 8.0
24:00 243.4 F 13.0
26:00 239.9 F 10.7
28:00 251.6 F 16.4
30:00 261.9 F 22.3
32:00 271.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.3
2:00 272.0 F 28.7
3:00 271.9 F 28.9
4:00 272.7 F 29.3

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 203.7 F 0.0
10:00 183.5 F 0.0
15:00 188.4 F 0.0
20:00 209.1 F 0.0
25:00 213.9 F 0.0
30:00 232.1 F 0.0

Drying Complete

FILLING: 0:42
HEATING: 32:40
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:11:02
CYCLE COMPLETE

01/29/2025 11:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____