

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002584  
7:16 01/11/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:58  
PRESSURIZING 2:05  
STERILIZING 2:58  
136.0 C 220kPa 2:58  
136.0 C 222kPa 3:28  
135.9 C 220kPa 3:58  
136.3 C 221kPa 4:28  
136.0 C 220kPa 4:58  
136.4 C 219kPa 5:28  
136.3 C 222kPa 5:58  
136.2 C 217kPa 6:28  
136.2 C 217kPa 6:28  
VENTING 6:28  
AIR DRYING 7:19  
CYCLE COMPLETE 7:21

-----  
Digital Signature #  
9CBFE97A70C74535  
-----