

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001381  
16:52 13/12/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:50  
PRESSURIZING 3:52  
STERILIZING 4:57  
136.0 C 221kPa 4:57  
136.3 C 224kPa 5:27  
136.4 C 224kPa 5:57  
136.4 C 223kPa 6:27  
136.2 C 223kPa 6:57  
136.4 C 225kPa 7:27  
136.4 C 224kPa 7:57  
136.4 C 224kPa 8:27  
136.4 C 224kPa 8:27  
VENTING 8:27  
AIR DRYING 9:21  
CYCLE COMPLETE 22:17  
-----

Digital Signature #  
729F4C83D19A0CB4  
-----