
Midmark M11 - v1.0.5
Total Cycles: 941

Name: RITTER2
Sterilizer ID: V2313228
04/01/2022 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.4 F 0.0
2:00 183.8 F 0.7
4:00 192.9 F 5.2
6:00 213.7 F 10.7
8:00 223.0 F 6.7
10:00 237.6 F 12.3
12:00 238.2 F 10.8
14:00 238.1 F 10.1
16:00 252.4 F 17.2
18:00 264.6 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.8 F 29.7
2:00 271.5 F 29.0
3:00 272.8 F 29.9
4:00 272.2 F 29.3

Min 271.4 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 212.1 F 0.0
10:00 195.2 F 0.0
15:00 196.7 F 0.0
20:00 200.8 F 0.0
25:00 219.8 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 19:53
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

04/01/2022 08:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____