

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3172

Name: RITTER 1  
Sterilizer ID: V2178198  
04/27/2023 04:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.5 F 0.0  
2:00 182.1 F 3.4  
4:00 199.7 F 7.9  
6:00 227.5 F 13.3  
8:00 229.7 F 8.4  
10:00 241.1 F 11.7  
12:00 241.0 F 11.5  
14:00 239.7 F 10.1  
16:00 253.6 F 17.1  
18:00 265.5 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.3 F 28.5  
2:00 272.6 F 29.0  
3:00 271.8 F 28.2  
4:00 272.8 F 28.9  
  
Min 271.6 F 28.1  
Max 272.9 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.0 F 0.0  
5:00 220.0 F 0.0  
10:00 232.6 F 0.0  
15:00 229.2 F 0.0  
20:00 229.2 F 0.0  
25:00 226.9 F 0.0  
30:00 225.0 F 0.0  
  
Drying Complete

FILLING: 1:36  
HEATING: 19:39  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:58:33  
CYCLE COMPLETE

04/27/2023 05:29 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_