
Midmark M11 - v1.0.0
Total Cycles: 2069

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2022 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.5 F 0.0
2:00 141.4 F 0.2
4:00 162.3 F 3.1
6:00 187.8 F 7.6
8:00 208.3 F 12.5
10:00 224.4 F 6.2
12:00 235.8 F 10.9
14:00 231.5 F 7.1
16:00 245.6 F 13.5
18:00 243.5 F 11.9
20:00 256.2 F 18.5
22:00 267.3 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.7 F 28.9
2:00 271.3 F 28.2
3:00 272.6 F 28.7
4:00 271.4 F 28.4

Min 271.3 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 210.9 F 0.0
10:00 216.1 F 0.0
15:00 216.2 F 0.0
20:00 258.0 F 0.0
25:00 234.7 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:41
CYCLE COMPLETE

06/28/2022 01:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____