
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006755
11:02 18/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 2:58
STERILIZING 3:57
136.4 C 222kPa 3:57
136.4 C 224kPa 4:27
136.4 C 224kPa 4:57
136.4 C 220kPa 5:27
136.4 C 223kPa 5:57
136.4 C 224kPa 6:27
136.2 C 224kPa 6:57
136.4 C 223kPa 7:27
136.4 C 223kPa 7:27
VENTING 7:27
AIR DRYING 8:24
CYCLE COMPLETE 10:37

Digital Signature #
1ADD73ECFFD5DE29
