

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001566  
10:35 07/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:56  
PRESSURIZING 2:59  
STERILIZING 3:56  
136.0 C 220kPa 3:56  
136.3 C 224kPa 4:26  
136.4 C 224kPa 4:56  
136.4 C 226kPa 5:26  
136.4 C 224kPa 5:56  
136.3 C 224kPa 6:26  
136.4 C 225kPa 6:56  
136.4 C 224kPa 7:26  
136.4 C 224kPa 7:26  
VENTING 7:26  
AIR DRYING 8:20  
CYCLE COMPLETE 8:51  
-----

Digital Signature #  
4D83C21CFEEBFB8D  
-----