
Midmark M11 - v1.0.0
Total Cycles: 3969

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2023 05:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.7 F 0.0
2:00 180.7 F 3.5
4:00 202.6 F 8.4
6:00 220.4 F 9.0
8:00 233.1 F 8.6
10:00 239.4 F 11.3
12:00 242.8 F 11.9
14:00 241.5 F 10.8
16:00 255.1 F 17.7
18:00 267.1 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.3 F 28.4
2:00 271.5 F 28.3
3:00 272.1 F 28.4
4:00 271.3 F 28.1

Min 271.2 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.0 F 0.0
5:00 227.4 F 0.0
10:00 226.7 F 0.0
15:00 224.0 F 0.0
20:00 204.5 F 0.0
25:00 200.8 F 0.0
30:00 206.3 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

11/01/2023 06:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____