
Midmark M11 - v1.0.5
Total Cycles: 1034

Name: RITTER2
Sterilizer ID: V2313228
04/26/2022 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.3 F 0.0
2:00 176.2 F 2.6
4:00 197.3 F 7.3
6:00 213.8 F 12.7
8:00 227.1 F 8.0
10:00 239.5 F 13.3
12:00 239.3 F 11.3
14:00 238.6 F 10.1
16:00 252.1 F 16.8
18:00 263.7 F 23.7
20:00 271.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.4 F 29.2
2:00 272.7 F 29.7
3:00 271.8 F 29.1
4:00 272.2 F 29.4

Min 271.7 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 226.3 F 0.0
10:00 245.1 F 0.0
15:00 228.3 F 0.0
20:00 223.2 F 0.0
25:00 226.0 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:57
HEATING: 20:02
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

04/26/2022 12:18 PM
ACCEPTED: X
REJECTED:
Operator: U02