
Midmark M11 - v1.0.5
Total Cycles: 391

Name: RITTER2
Sterilizer ID: V2313228
09/23/2021 12:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.9 F 0.0
2:00 175.9 F 3.7
4:00 204.6 F 9.4
6:00 217.5 F 6.8
8:00 233.3 F 11.6
10:00 235.7 F 9.7
12:00 236.8 F 9.1
14:00 250.4 F 16.0
16:00 262.7 F 23.2
18:00 272.1 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.4 F 29.4
2:00 272.8 F 29.9
3:00 271.7 F 29.2
4:00 272.4 F 29.5

Min 271.6 F 28.9
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 224.7 F 0.0
10:00 227.1 F 0.0
15:00 226.7 F 0.0
20:00 223.5 F 0.0
25:00 222.5 F 0.0
30:00 220.6 F 0.0

Drying Complete

FILLING: 2:51
HEATING: 18:07
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

09/23/2021 01:10 PM
ACCEPTED: X
REJECTED:
Operator: U02