

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1897

Name: RITTER 1  
Sterilizer ID: V2178198  
05/02/2022 03:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
2:00 194.7 F 3.8  
4:00 209.7 F 8.4  
6:00 225.0 F 9.2  
8:00 233.6 F 10.0  
10:00 236.9 F 8.2  
12:00 244.3 F 12.3  
14:00 245.0 F 12.8  
16:00 257.9 F 19.8  
18:00 269.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.1 F 28.8  
2:00 271.9 F 29.1  
3:00 272.0 F 28.6  
4:00 272.2 F 29.3  
  
Min 271.4 F 28.4  
Max 272.4 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
5:00 225.5 F 0.0  
10:00 231.4 F 0.0  
15:00 228.6 F 0.0  
20:00 228.3 F 0.0  
25:00 230.6 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 4:42  
HEATING: 18:57  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:51  
CYCLE COMPLETE

05/02/2022 04:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_