
Midmark M11 - v1.0.0
Total Cycles: 941

Name: RITTER 1
Sterilizer ID: V2178198
06/25/2021 06:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.2 F 0.0
2:00 76.4 F 0.0
4:00 89.9 F 0.0
6:00 109.8 F 0.0
8:00 153.0 F 0.1
10:00 186.1 F 0.3
12:00 208.0 F 1.5
14:00 221.8 F 4.5
16:00 229.1 F 7.2
18:00 236.8 F 10.4
20:00 240.1 F 9.8
22:00 241.0 F 11.1
24:00 236.4 F 8.5
26:00 241.9 F 10.6
28:00 247.3 F 13.4
30:00 258.8 F 19.6
32:00 269.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.8
1:00 272.6 F 28.5
2:00 271.6 F 27.7
3:00 272.5 F 28.6
4:00 272.2 F 28.2

Min 271.4 F 27.7
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 228.6 F 0.0
10:00 227.0 F 0.0
15:00 227.5 F 0.0
20:00 228.4 F 0.0
25:00 232.9 F 0.0
30:00 232.9 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 33:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:11:14
CYCLE COMPLETE

06/25/2021 08:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____