

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1121

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2022 09:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.8 F 0.0  
2:00 161.7 F 2.8  
4:00 193.8 F 7.2  
6:00 213.8 F 12.7  
8:00 227.0 F 7.8  
10:00 236.8 F 13.0  
12:00 238.5 F 10.4  
14:00 236.4 F 8.9  
16:00 250.6 F 15.9  
18:00 262.3 F 22.7  
20:00 272.9 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 271.7 F 29.3  
2:00 272.0 F 29.3  
3:00 272.5 F 29.4  
4:00 272.1 F 29.6

Min 271.4 F 29.2  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 239.0 F 0.0  
10:00 215.4 F 0.0  
15:00 218.9 F 0.0  
20:00 179.0 F 0.0  
25:00 179.0 F 0.0  
30:00 189.8 F 0.0  
Drying Complete

FILLING: 2:08  
HEATING: 20:59  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 01:00:16  
CYCLE COMPLETE

05/26/2022 11:13 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02