
Midmark M11 - v1.0.0
Total Cycles: 4143

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2023 04:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 172.3 F 3.8
4:00 193.5 F 8.4
6:00 214.1 F 13.5
8:00 227.5 F 8.3
10:00 238.3 F 12.3
12:00 240.9 F 11.0
14:00 238.1 F 9.3
16:00 252.0 F 16.0
18:00 263.8 F 22.8
20:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 271.9 F 28.2
2:00 272.4 F 28.6
3:00 271.3 F 28.2
4:00 272.3 F 28.3

Min 271.2 F 27.9
Max 272.6 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.2 F 0.0
5:00 196.7 F 0.0
10:00 195.5 F 0.0
15:00 191.9 F 0.0
20:00 188.9 F 0.0
25:00 186.6 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 20:01
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

12/20/2023 05:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____