
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001579
17:06 08/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:12
STERILIZING 3:11
136.0 C 223kPa 3:11
136.1 C 223kPa 3:41
136.4 C 226kPa 4:11
136.1 C 220kPa 4:41
136.4 C 226kPa 5:11
136.4 C 226kPa 5:41
136.2 C 224kPa 6:11
136.4 C 226kPa 6:41
136.4 C 226kPa 6:41
VENTING 6:41
AIR DRYING 7:33
CYCLE COMPLETE 7:39

Digital Signature #
D38581FF52ECD099
