
Midmark M11 - v1.0.5
Total Cycles: 80

Name: RITTER2
Sterilizer ID: V2313228
06/09/2021 07:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 74.9 F 0.2
4:00 79.6 F 1.3
6:00 105.3 F 3.5
8:00 139.7 F 6.0
10:00 166.1 F 9.1
12:00 189.0 F 12.6
14:00 198.9 F 4.4
16:00 212.9 F 8.1
18:00 225.2 F 12.2
20:00 226.9 F 8.2
22:00 239.0 F 13.3
24:00 239.9 F 11.5
26:00 251.1 F 17.4
28:00 261.5 F 23.4
30:00 271.4 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.0
1:00 272.8 F 30.6
2:00 273.2 F 30.4
3:00 272.0 F 30.0
4:00 273.0 F 30.8

Min 271.8 F 29.8
Max 273.5 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.5 F 0.0
5:00 231.1 F 0.0
10:00 227.8 F 0.0
15:00 226.6 F 0.0
20:00 226.4 F 0.0
25:00 227.4 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:32
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:37
CYCLE COMPLETE

06/09/2021 08:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____